

Notice

- Please add the liquid before turning on the device.
- Please do not use health products containing oil or liquid atomization.
- Please follow doctor's advice on drug selection, drug usage and treatment time.
- Please do not use distilled water, multiple-filtered purified water or oil substances for nebulization. Please dean and maintain the device after each

- Please keep the device, medication cup and accessories dry after each time use.

 Please do not touch the mesh with sharp or hard objects or with figure in case of malfunction.
- Please charge at least 30 minutes before first use.





Operating Steps



Open the medication cup lid



Add proper liquid drug. Please do not exceed the max scale indication.




Close the medication cup lid




Connect mask or mouthpiece




Start nebulizing







Note:There is enough liquid, but the orange light flashes ten times,during orange light flashing please press switch key once to continue work.



Note:Before charging, please make sure the ∆ socket connected is normally charged.

A Note: This instrument is independent charging, please do not charging with any other electronic equipment.



(Mesh Diaphragm Operating Steps)

Rotate counterclockwise to remove the mesh diaphragm, replace the mesh diaphragm, rotate clockwise, and install.

Note:When installing the mesh diaphragm, the metal probe should be in the arc-shaped groove area of the host, and the two sides of the mesh diaphragm should be connected to the host card slot. When a "click" sound is heard, installation is completed.

Note: please keep all contact points clean and dry after use.

Cleaning Guidance 1. Preliminary Cleaning



Rotate the accessories clockwise so as to loose them and then pull them out.



Open the medication cup cover and clean the drug liquid residue.



Add appropriate amount of warm water (Below 40 $^{\circ}$ C). Please do not use distilled water or multiple-filtered purified water



Close the medication cup lid



Please wipe the machine with clean medical gauze after nebulization cleaning. Do not touch the central area of the mesh to avoid the damage of the mesh.

2. Regular Cleaning



Hold down the press key and pull up





Please wash the external and internal of the medication cup and spraying connecting interface repeatedly with purified water(Below 40°C). Please dry the liquid drops with clean medical gauze. Please do not touch the center of the mesh in case of damage. Please put the device and accessories back to the given place and store them as instructed in the manual. instructed in the manual.

Tip: Nebulization with 3 drops purified water and 1 drop white vinegar will achieve better cleaning effect.

Notice: Diagram is only for reference. Alternation of

Notice: Diagram is only for reference. Alternation of device color is without further notice. Detailed information in only according to actual product.